

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 21 2009


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130183
Name		HUTT, William T.	ID# 3359 Date 05/14/09
A	Agency	Homer Police Dept. / Jail	Phone # 235-3150
			235-3151 (fax)
Instrument Location		4060 Heath St. Homer, AK 99603	
Alco S/N		X301599	Target Value .086
			High Pressure 1100 psi
B	Alco Test Values	.082	.082
		1 st Alco	2 nd Alco
Signature		William T. Hutt	
(OVER)			

BMB
5/22/09


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 27th day of May 2009.


Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 138103 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

MAY 14, 2009

OPERATOR'S NAME:

NOTTY WILSON

OPERATOR'S NUMBER: 3359

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

ETOH

O.L. #: 01234567

DEPT/AGENCY: HOM1

CASE/REPORT: 09-0000

TEST TYPE: U

ALCO TARGET VALUE: .086 ✓

ALCO S/N: X301599 ✓

— BREATH ANALYSIS —

.086 ADJUSTED FOR 29.85 in ✓
ALCO TARGET .085 15:46
BLANK TEST .000 15:47
INTERNAL STANDARD VERIFIED 15:47
ALCO TV 29.85 in .082 ✓ 15:48
BLANK TEST .000 15:48
SUBJECT SAMPLE WTHR 15:49
BLANK TEST .000 15:50
ALCO TV 29.85 in .082 ✓ 15:51
BLANK TEST .000 15:51

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130183

MAY 14, 2009 ✓

TIME 15:43

WTHR

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
REUTERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.85 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■